

**VAPOR-TRAK®** Xylene Monitors  
**Laboratory Analysis Report**



**Kem Laboratory:** 5305 NW 35th Terrace  
 Ft Lauderdale FL 33309  
 (954) 733-7499 • (800) 875-9028  
 Fax: (954) 733-9908

400 Broadhollow Road, Ste 2  
 Farmingdale NY 11735  
 (631) 454-6565 • (800) 553-0330  
 Fax: (631) 454-8083

**Direct all questions to:**  
**(800) 875-9028 or (954) 733-7499**

**Customer ID #: X3337**

**Authorization Contact:** D SMITH

**Facility:** XYZ HOSPITAL

**Address:** 123 MAIN ST  
 ANYTOWN NY 10000

**Phone:** 123 456-7890

**Xylene  
 Federal Standard**  
 100 ppm - 8 hour TWA - PEL  
 150 ppm - 15 minute TWA - STEL

Analysis ID #: X160430001

Vapor Analyzed: **Xylene**

Monitor Type: Passive Dosimeter Date Received: 04/30/16 Date Analyzed: 04/30/16 Monitor condition: Satisfactory

Person or Area monitored	Badge number	Sampling date (mm/dd/yy)	Mass recovered (µg)	Sampling time (Hrs.)	Exposure concentration (ppm - TWA)
HISTOLOGY	XX5551	04/25/16	2.00	8.00	0.02

**Remark:**  
 COVERSLEPPING

Exposure Profile™ Monitoring History As of 4/30/16	Number of samples on record for the person/area monitored	Highest exposure concentration (ppm - TWA)	Lowest exposure concentration (ppm - TWA)	Average exposure concentration (ppm - TWA)
	1	0.02	0.02	0.02

**Lowest detectable limit using this method is:** 0.02 ppm

**Overall System Accuracy (OSA)** ±5.65%

**Method of Analysis:** Modified NIOSH method #1501

The molar volume at 25°C (24.45 L/mol) was used to calculate ppm. Air concentrations are based upon field sampling information provided by the customer.

Laboratory Chain of Custody

Received by: GM  
 Analyst: DM  
 Analysis Approval: CW

**Employee Review:**

X \_\_\_\_\_ Date: \_\_\_\_\_

I have seen and reviewed the results of my monitoring on the date above.

Report By: D. Maurer  
 D. Maurer, Lab Director

All applicable QC were within method specifications any customer supplied field blanks were not subtracted from sample results unless otherwise noted, the following statements apply:

1) All samples were received in acceptable condition; 2) all Quality Control results associated with this sample set were within acceptable limits and/or do not adversely affect the reported results, and 3) the results relate only to items tested.

This report is not to be reproduced except in full, without written permission of Kem Medical Products.



9001:2008  
 Certification #2102US